

# FOR EMERGENCY USE!



Cochlear Implant Users need **accessibility to communication and Cochlear Implant Care & help during hospitalization**

NAME: \_\_\_\_\_

DATE OF BIRTH: \_\_\_\_\_

ADDRESS: \_\_\_\_\_

MY CONTACT FOR QUESTIONS: \_\_\_\_\_

## I HAVE HEARING LOSS AND WEAR THESE DEVICES:

**Left:** Cochlear Implant / Hearing Aid

Model and Brand: \_\_\_\_\_

**Right:** Cochlear Implant / Hearing Aid

Model and Brand: \_\_\_\_\_

**Important:** Do not place devices on the wrong ears!



Cochlear implants (CI) replace natural hearing with electronic hearing. Keep in mind:

I cannot hear anything without my external Cochlear Implant (CI) processor(s).

Look at me! Eye contact and the possibility to lip-read will help us to communicate.

**CI processor on and off:** Insert and remove batteries or press the top button.


**To attach the processor:** Connect the coil of the external processor to the invisible magnet under the scalp (located slightly above / behind the ear).

**To replace disposable batteries:** Slide the battery cover downwards to expose the batteries. Insert two zinc-air 675P batteries with the flat side facing upwards.

**Batteries usually last:** \_\_\_\_\_



© Oticon Medical



© Oticon Medical

At night, place the cochlear implant processor(s) in the dehumidifier if provided.

Please do not touch the remote control(s).

**DO NOT** perform MRI without prior clarification and professional agreement with the otolaryngologist at the cochlear implantation clinic and the **manufacturer**:

**If necessary, please write down important messages clearly so that I can read them!**

© EURO-CIU, asbl [www.eurociu.eu](http://www.eurociu.eu)

© Bayerischer Cochlea-Implantat-Verband e.V., [www.bayciv.de](http://www.bayciv.de)